

**Item Name :** Volumetric Flask One Graduation Mark

**Item Code :** 140.2008.01

**Item Description :**

With Polythelene stopper ISO 1042, DIN 12664, Class B

**Item Image :**



Cat. No.	Cap. ml.	Tolerance ( $\pm$ ml)	N/S
140.2008.01	5	0.05	10/19
140.2008.02	10	0.05	10/19
140.2008.02A	20	0.08	10/19
140.2008.03	25	0.08	10/19
140.2008.04	50	0.12	12/21
140.2008.05	100	0.20	14/23
140.2008.06	200	0.30	14/23
140.2008.07	250	0.30	14/23
140.2008.08	500	0.50	19/26
140.2008.09	1000	0.80	19/26

140.2008.10	2000	1.20	29/32
140.2008.11	5000	2.40	34/35

# 44, Vikas puri, P.O.Ind.Estate, Jagadhri Road,  
A/Cantt-133 006. Haryana, India.

+0171 – 2698107

[info@glacierindia.com](mailto:info@glacierindia.com)

<https://www.glacierindia.com>